



Government of Gujarat  
**L. D. College of Engineering, Ahmedabad**  
Opp. Gujarat University, Navrangpura  
Ahmeabad - 380 015

Phone : Office - 079 26306752, Principal - 079 26302887

Email : ldce-abad-dte@gujarat.gov.in Website : www.ldce.ac.in



LDCE/Chemical/Equip Repairing/ 254

Date:- 22-01-2021

## INQUIRIES

To,

As per list

Sub : Quotation for equipment repairing

Sir,

The Quotations are invited for the repairing of equipments of Chemical Engineering Department, L D College of Engineering, Ahmedabad. List of the same is enclosed herewith. Terms and conditions are:

1. The quotation should be sent in a sealed cover duly superscripted as **“Quotation for Equipment Repairing- Chemical Engineering” to The Principal, L. D. College of Engineering, Ahmedabad.**
2. The last date for receiving the quotation is 12/2/2021
3. The details of repairing of each item are attached as per the annexure attached herewith.
4. Rates must be quoted for repairing of the each equipment separately.
5. Your rates should be F.O.R. should be inclusive of all taxes.
6. The validity period for the quotation should be 3 months from the date on which the tender is due.
7. Repairing work must be fulfilled within 60 days of time period from the date of work order.
8. Payment will be released only after inspection and satisfactory approval of repairing work.
9. This office reserves the right to reject any or all quotations without assigning any reasons.

  
PRINCIPAL

As per sheet on page No.2.

### Annexure

Sr no	Equipment Name	Description
1	Centrifugal pump test rig(variable speed type)	DC drive, 6 amp toggle switch, Top 15 amp., Repairing & Servicing
2	Aniline point app.	voltage regulator, Coil, Wiring, Repairing & Servicing
3	Distillation app.	Coil, Top 6 amp., Wiring , Repairing & Servicing
4	Thermocouple calibration test rig	Temperature indicator, thermocouples all type, Wiring , Repairing & Servicing
5	Level control Trainer	6 amp toggle switch, lamp, pump, LLC SPZP circuit, Wiring, Repairing & Servicing
6	Cascade Continuous stirred reactor	Ball valves, DC drive, PID, 5 ampere toggle switch, lamp, solenoid valve 1/4 horse collar, Air regulator, pressure gauge, 50 to 300 C thermostat, wiring and service
7	BOD	Controller, Wiring ,Repairing & Servicing
8	Muffle furnace	ROPE new. Wiring , Repairing & Servicing
9	Cooling tower	PID, SSR, heater, 6 zone indicator, RTD Sensor, Valve. Top 15 amp., Repairing & Servicing
10	Batch crystallizer	PID, heater, Wiring ,Repairing & Servicing
11	Sieve shaker	sieves, Wiring, Repairing & Servicing
12	Heat transfer in forced convection	12 zone indicator, J type sensor, Wiring, Repairing & Servicing